

Certificate of Quality according to ISO 4074:2002(E) for Male Latex Condoms

Client:	RRT VERTRIEB UND SERVICE	Rrt Shipment No.:	SO/000014-01-17	Order No.:	V&S 004/17
Brand Name:	IN LOVE FORMATO AROMAS	Quantity:	95 GROSS	Lubrication:	SILICONE OIL + COLA FLAVOUR
RRT Batch No.:	PL7099C	Manufacturing Date:	2017 - 01	Sealed:	(55 X 55)mm
Type/Size :	PLAIN 53mm	Expiry Date:	2021 - 12	Colour:	BROWN

1. Dimensions:

Sample No.:	1	2	3	4	5	6	7	8	9	10	11	12	13	Std(mm)	Average(mm)	Results
Length (mm)	192	195	193	194	192	193	195	191	192	193	194	195	193	≥ 190	193	Pass
Width (mm)	53	53	53	53	53	53	53	53	53.5	53.5	53	53	53	53 ± 2	53.1	Pass
Thickness (mm)	0.073	0.069	0.071	0.072	0.073	0.072	0.071	0.073	0.072	0.071	0.069	0.073	0.071	0.065±0.015	0.072	Pass

2. a) Burst Volume (Unaged) : General Inspection Level 1, AQL 1.5

b) Burst Pressure : General Inspection Level 1, AQL 1.5

Sample Nos.:	Accept - Reject	Defect	Min. Volume 18dm ³	Result	Accept - Reject	Defect	Min. Pressure 1.0 kPa	Result
200	7 8	0	Mean :34.9	Pass	7 8	0	Mean :2.00	Pass

3. Lubricant Quantity : Inspection Level S-2, AQL 4.0

Sample Nos.:	Accept - Reject	1	2	3	4	5	6	7	8	9	10	11	12	13	Std (mg)	Average(mg)	Results
13	1 2	389	377	381	376	369	387	390	377	369	382	376	388	391	350 - 400	381	Pass

4. Freedom from Holes :General Inspection Level 1, AQL 0.25,Code Letter M

Sample Nos.:	Accept - Reject	Defect	Result
315	2 3	0	Pass

5. Package Integrity, Special Inspection Level S-3, AQL 2.5

Sample Nos.:	Accept - Reject	Defect	Result
32	2 3	0	Pass

6. Visible defects: General Inspection Level 1, AQL 0.4


Sample Nos.:	Accept - Reject	Defect	Result
315	3 4	0	Pass

7. Packaging and Labelling, S-3, AQL 2.5

Sample Nos.:	Accept - Reject	Defect	Result
13	0 1	0	Pass


Result: pass / reject

Prepared by:
QA Executive



Date: 06/03/17

Verified by:
QA Manager



Date:

06.03.2017