

Certificates of Analysis according to ISO 4074:2015 (E) for Male Latex Condom

Customer:	RRT VERTRIEB UND SERVICE	Shipment No.:	SO/000153-06-21	RRT Order No.:	V&S005/21
Brand Name:	IN LOVE MIX PROFESIONAL 144'S	Lot Size:	50 GROSS	Sealing:	55 x 55 MM
RRT Lot No.:	PL1538N	Manufacturing Date:	2021/06	Lubrication:	SILICONE OIL
Type / Size :	PLAIN 57 MM	Expiry Date:	2026/05	Colour:	NATURAL
Product family:	STANDARD CONDOM				

1. Dimensions: All shall comply

Sample No.:	1	2	3	4	5	6	7	8	9	10	11	12	13	Standard (mm)	Average (mm)	Results
Length (mm), Annex D	192	193	193	194	193	192	193	194	191	192	193	193	192	≥ 190	193	Pass
Width (mm), Annex E	57.0	57.0	57.0	57.0	57.0	57.0	57.0	56.5	57.0	57.0	57.0	57.0	57.0	57 ± 2	57.0	Pass
Thickness (mm), Annex F	0.069	0.070	0.069	0.068	0.067	0.067	0.068	0.068	0.067	0.069	0.068	0.069	0.068	0.068 ± 0.010	0.068	Pass

2. a) Burst Volume (Unaged): General Inspection Level 1, AQL 1.5, Annex H

Sample Nos.:	Accept - Reject	Defect	Min. Volume (dm³)	22 dm³	Result
200 pcs	7 - 8	1	Mean : 38.2		Pass

2. b) Burst Pressure: General Inspection Level 1, AQL 1.5, Annex H

Sample Nos.:	Accept - Reject	Defect	Min. Pressure: 1.0 kPa	Result
200 pcs	7 - 8	0	Mean: 1.84	Pass

3. a) Burst Volume (Aged): General Inspection Level 1, AQL 1.5, Annex G, ISO 4074:2002 (E)

Sample Nos.:	Accept - Reject	Defect	Min. Volume (dm³)	NA	Result
NA	NA	NA	NA	NA	NA

3. b) Burst Pressure: General Inspection Level 1, AQL 1.5, Annex G, ISO 4074:2002 (E)

Sample Nos.:	Accept - Reject	Defect	Min. Pressure: NA	Result
NA	NA	NA	NA	NA

4. a) Freedom from Holes: General Inspection Level 1, AQL 0.25, Code Letter M, Annex M

Sample Nos.:	Accept - Reject	Defect	Result
315 pcs	2 - 3	0	Pass

4. b) Visibly open seal: General Inspection Level 1, AQL 0.4, Code Letter M, Annex M

Sample Nos.:	Accept - Reject	Defect	Result
315 pcs	3 - 4	0	Pass

5. Lubricant Quantity: All shall comply, Annex C

Sample Nos.:	1	2	3	4	5	6	7	8	9	10	11	12	13	Acc - Rej	Std (mg)	Average (mg)	Results
13 pcs	398	388	389	390	373	376	368	363	371	393	387	394	390	0 - 1	350 - 400	383	Pass

6. Package Integrity, Special Inspection Level S-3, AQL 2.5, Annex N

Sample Nos.:	Accept - Reject	Defect	Result
32 pcs	2 - 3	0	Pass

7. Visible defects: General Inspection Level 1, AQL 0.4, Annex M

Sample Nos.:	Accept - Reject	Defect	Result
315 pcs	3 - 4	0	Pass

8. Packaging and Labelling, All shall comply

Sample Nos.:	Accept - Reject	Defect	Result
13 pcs	0 - 1	0	Pass

Status: Approved

Prepared by:
QA Sr. Supervisor

Date: 31.12.21

Checked by:
QA Executive

Date: 04.01.2022

Approved by:
QA Manager/Designee

Date: 12.01.2022